



## MEGHALAYA BASIN MANAGEMENT AGENCY (MBMA)

Regd. Address: House No. L/ A-56 Lower Nongrim Hills,

Shillong, East Khasi Hills, Meghalaya- 793003

(CIN No. U75144ML2012NPL008509)

Phone: +91-364- 2522921/2522992

Website: [www.mbda.gov.in](http://www.mbda.gov.in), E-mail: [admin.mbda@gov.in](mailto:admin.mbda@gov.in)

Issued on: March 25, 2023

### Clarification, Addendum and Corrigendum no. 1:

**“Procurement of 100 Nos. of Automatic Weather Stations for the office of Meghalaya Basin Management Agency (MBMA)”**  
**RFB Bid No: MBMA/CLLMP/2022-23/GD-74, Date of Issue of Bid Document: March 06, 2023**

Further to the RFB, the clarification, amendment, corrigendum, notified as below. All other requirements information, terms and conditions shall remain unchanged in the RFB

#### 1. Clarification of the pre-bid queries:

Sl. No.	Clause Ref.	Original Clause	MBMA response
1.	Page No. 65-66 Requirements	A 3-meter-high triangular guy mast must be supplied, installed, and commissioned	A 3-meter height of SS Tripod with the self-supporting structure and lightning arrestor should be provided
2.	Page No. 65-66 Requirements	Industrial grade spiral fences of high strength aluminum alloy with corrosive resistance and fastened with fiber reinforced Plastic FRP screws and entry gate with high strength aluminum alloy with corrosive resistance	GI chain link fencing of dimensions 5m x 5m and height shall be 2 meters
3.	Page No. 68 NETWORK MANAGEMENT SOFTWARE:	A Network Management Software (NMS) shall be provided by the vendor to maintain individual station wise data which are connected with NMS through GSM/GPRS	The vendor should provide the NMS software and MBMA will provide the server/hardware
4.	Page No.73 BATTERY & CHARGING SYSTEM	2. Bidder Should provide Sealed Maintenance Free Lead Acid (VRLA) Batteries for 12V/75 Amp-Hour capacity or higher. 4. The combined capacity of the Solar panels should be 150Watts for 12 VDC configuration.	Bidder shall quote the power supply and provide the power budget calculation and shall consider 20 days battery economy during cloudy day or without charging the battery
5.	Page Number 41 - Section 3.1	The bidder must have successfully completed project	There is no restriction in the number of work orders but